



December 12, 2024

Don Virgilio
NYS School for the Blind
2A Richmond Avenue
Batavia, NY 14020

RE: Project: LEAD TESTING 11/16
Pace Project No.: 70326093

Dear Don Virgilio:

Enclosed are the analytical results for sample(s) received by the laboratory on December 06, 2024. The results relate only to the samples included in this report. Results reported herein conform to the applicable TNI/NELAC Standards and the laboratory's Quality Manual, where applicable, unless otherwise noted in the body of the report.

The test results provided in this final report were generated by each of the following laboratories within the Pace Network:

- Pace Analytical Services - Melville

If you have any questions concerning this report, please feel free to contact me.

Sincerely,

A handwritten signature in cursive script that reads "Daniel H. Bonitto".

Daniel H. Bonitto
daniel.bonitto@pacelabs.com
516-370-6000
Project Manager

Enclosures



REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



CERTIFICATIONS

Project: LEAD TESTING 11/16

Pace Project No.: 70326093

Pace Analytical Services, LLC - Melville, NY

575 Broad Hollow Rd, Melville, NY 11747

Connecticut Certification #: PH-0435

Delaware Certification # NY 10478

Maryland Certification #: 208

Massachusetts Certification #: M-NY026

New Hampshire Certification #: 2987

New Jersey Certification #: NY158

New York Certification #: 10478 Primary Accrediting Body

Pennsylvania Certification #: 68-00350

Rhode Island Certification #: LAO00340

Texas Certification #: T104704582

Florida Certification #: E871198

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: LEAD TESTING 11/16

Pace Project No.: 70326093

Sample: K1 STEAM KETTLE SPOUT Lab ID: 70326093001 Collected: 11/16/24 06:11 Received: 12/06/24 08:00 Matrix: Drinking Water

Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water	Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	<1.0	ug/L	1.0	1		12/11/24 11:11	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: LEAD TESTING 11/16

Pace Project No.: 70326093

Sample: **K5 KITCHEN COMP SINK RIGHT** Lab ID: **70326093002** Collected: 11/16/24 06:14 Received: 12/06/24 08:00 Matrix: Drinking Water

Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water	Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	3.6	ug/L	1.0	1		12/11/24 11:12	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: LEAD TESTING 11/16

Pace Project No.: 70326093

Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
Sample: B131 STUDENT LOUNGE Lab ID: 70326093003 Collected: 11/16/24 06:17 Received: 12/06/24 08:00 Matrix: Drinking Water								
200.8 MET ICPMS Drinking Water	Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	17.6	ug/L	1.0	1		12/11/24 11:19	7439-92-1	M1

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: LEAD TESTING 11/16

Pace Project No.: 70326093

Sample: C127 SOUTH COTTAGE KITCHEN **Lab ID: 70326093004** Collected: 11/29/24 06:15 Received: 12/06/24 08:00 Matrix: Drinking Water

Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water	Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	2.2	ug/L	1.0	1		12/11/24 11:24	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: LEAD TESTING 11/16

Pace Project No.: 70326093

Sample: **D247B HEALTH CENTER CLINIC 1** Lab ID: **70326093005** Collected: 11/29/24 06:24 Received: 12/06/24 08:00 Matrix: Drinking Water

Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water	Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	8.4	ug/L	1.0	1		12/11/24 11:28	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: LEAD TESTING 11/16

Pace Project No.: 70326093

Sample: **D226 HEALTH CENTER CLINIC 2** Lab ID: **70326093006** Collected: 11/29/24 06:20 Received: 12/06/24 08:00 Matrix: Drinking Water

Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water	Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	74.7	ug/L	1.0	1		12/11/24 11:30	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: LEAD TESTING 11/16

Pace Project No.: 70326093

Sample: **D227 HEALTH CENTER CLINIC 2** Lab ID: **70326093007** Collected: 11/29/24 06:22 Received: 12/06/24 08:00 Matrix: Drinking Water

Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water	Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	118	ug/L	1.0	1		12/11/24 11:34	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: LEAD TESTING 11/16

Pace Project No.: 70326093

Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
Sample: D242 KITCHEN								
Lab ID: 70326093008								
Collected: 11/29/24 06:26								
Received: 12/06/24 08:00								
Matrix: Drinking Water								
200.8 MET ICPMS Drinking Water								
Analytical Method: EPA 200.8								
Pace Analytical Services - Melville								
Lead	9.4	ug/L	1.0	1		12/11/24 11:36	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full, without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: LEAD TESTING 11/16

Pace Project No.: 70326093

Sample: D242 F KITCHEN		Lab ID: 70326093009	Collected: 11/29/24 06:27	Received: 12/06/24 08:00	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	<1.0	ug/L	1.0	1		12/11/24 11:37	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



QUALITY CONTROL DATA

Project: LEAD TESTING 11/16

Pace Project No.: 70326093

QC Batch: 374792	Analysis Method: EPA 200.8
QC Batch Method: EPA 200.8	Analysis Description: 200.8 MET No Prep Drinking Water
	Laboratory: Pace Analytical Services - Melville

Associated Lab Samples: 70326093001, 70326093002

METHOD BLANK: 1963281 Matrix: Water

Associated Lab Samples: 70326093001, 70326093002

Parameter	Units	Blank Result	Reporting Limit	Analyzed	Qualifiers
Lead	ug/L	<1.0	1.0	12/11/24 10:28	

LABORATORY CONTROL SAMPLE: 1963282

Parameter	Units	Spike Conc.	LCS Result	LCS % Rec	% Rec Limits	Qualifiers
Lead	ug/L	50	48.1	96	85-115	

MATRIX SPIKE SAMPLE: 1963285

Parameter	Units	70326370002 Result	Spike Conc.	MS Result	MS % Rec	% Rec Limits	Qualifiers
Lead	ug/L	<1.0	25	33.5	132	70-130	M1

MATRIX SPIKE SAMPLE: 1963287

Parameter	Units	70326371002 Result	Spike Conc.	MS Result	MS % Rec	% Rec Limits	Qualifiers
Lead	ug/L	9.8	25	40.6	123	70-130	

SAMPLE DUPLICATE: 1963284

Parameter	Units	70326370002 Result	Dup Result	RPD	Qualifiers
Lead	ug/L	<1.0	<1.0		

SAMPLE DUPLICATE: 1963286

Parameter	Units	70326371002 Result	Dup Result	RPD	Qualifiers
Lead	ug/L	9.8	10.0	3	

Results presented on this page are in the units indicated by the "Units" column except where an alternate unit is presented to the right of the result.

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full, without the written consent of Pace Analytical Services, LLC.



QUALITY CONTROL DATA

Project: LEAD TESTING 11/16

Pace Project No.: 70326093

QC Batch:	374793	Analysis Method:	EPA 200.8
QC Batch Method:	EPA 200.8	Analysis Description:	200.8 MET No Prep Drinking Water
		Laboratory:	Pace Analytical Services - Melville
Associated Lab Samples:	70326093003, 70326093004, 70326093005, 70326093006, 70326093007, 70326093008, 70326093009		

METHOD BLANK:	1963288	Matrix:	Water
Associated Lab Samples:	70326093003, 70326093004, 70326093005, 70326093006, 70326093007, 70326093008, 70326093009		

Parameter	Units	Blank Result	Reporting Limit	Analyzed	Qualifiers
Lead	ug/L	<1.0	1.0	12/11/24 11:17	

LABORATORY CONTROL SAMPLE: 1963289						
Parameter	Units	Spike Conc.	LCS Result	LCS % Rec	% Rec Limits	Qualifiers
Lead	ug/L	50	48.6	97	85-115	

MATRIX SPIKE SAMPLE: 1963291							
Parameter	Units	70326093003 Result	Spike Conc.	MS Result	MS % Rec	% Rec Limits	Qualifiers
Lead	ug/L	17.6	25	50.3	131	70-130	M1

MATRIX SPIKE SAMPLE: 1963293							
Parameter	Units	70326093004 Result	Spike Conc.	MS Result	MS % Rec	% Rec Limits	Qualifiers
Lead	ug/L	2.2	25	33.2	124	70-130	

SAMPLE DUPLICATE: 1963290					
Parameter	Units	70326093003 Result	Dup Result	RPD	Qualifiers
Lead	ug/L	17.6	17.4	1	

SAMPLE DUPLICATE: 1963292					
Parameter	Units	70326093004 Result	Dup Result	RPD	Qualifiers
Lead	ug/L	2.2	2.3	4	

Results presented on this page are in the units indicated by the "Units" column except where an alternate unit is presented to the right of the result.

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full, without the written consent of Pace Analytical Services, LLC.



QUALIFIERS

Project: LEAD TESTING 11/16

Pace Project No.: 70326093

DEFINITIONS

DF - Dilution Factor, if reported, represents the factor applied to the reported data due to dilution of the sample aliquot.

ND - Not Detected at or above adjusted reporting limit.

TNTC - Too Numerous To Count

J - Estimated concentration above the adjusted method detection limit and below the adjusted reporting limit.

MDL - Adjusted Method Detection Limit.

PQL - Practical Quantitation Limit.

RL - Reporting Limit - The lowest concentration value that meets project requirements for quantitative data with known precision and bias for a specific analyte in a specific matrix.

S - Surrogate

1,2-Diphenylhydrazine decomposes to and cannot be separated from Azobenzene using Method 8270. The result for each analyte is a combined concentration.

Consistent with EPA guidelines, unrounded data are displayed and have been used to calculate % recovery and RPD values.

LCS(D) - Laboratory Control Sample (Duplicate)

MS(D) - Matrix Spike (Duplicate)

DUP - Sample Duplicate

RPD - Relative Percent Difference

NC - Not Calculable.

SG - Silica Gel - Clean-Up

U - Indicates the compound was analyzed for, but not detected.

N-Nitrosodiphenylamine decomposes and cannot be separated from Diphenylamine using Method 8270. The result reported for each analyte is a combined concentration.

Reported results are not rounded until the final step prior to reporting. Therefore, calculated parameters that are typically reported as "Total" may vary slightly from the sum of the reported component parameters.

Pace Analytical is TNI accredited. Contact your Pace PM for the current list of accredited analytes.

TNI - The NELAC Institute.

ANALYTE QUALIFIERS

M1 Matrix spike recovery exceeded QC limits. Batch accepted based on laboratory control sample (LCS) recovery.

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



QUALITY CONTROL DATA CROSS REFERENCE TABLE

Project: LEAD TESTING 11/16

Pace Project No.: 70326093

Lab ID	Sample ID	QC Batch Method	QC Batch	Analytical Method	Analytical Batch
70326093001	K1 STEAM KETTLE SPOUT	EPA 200.8	374792		
70326093002	K5 KITCHEN COMP SINK RIGHT	EPA 200.8	374792		
70326093003	B131 STUDENT LOUNGE	EPA 200.8	374793		
70326093004	C127 SOUTH COTTAGE KITCHEN	EPA 200.8	374793		
70326093005	D247B HEALTH CENTER CLINIC 1	EPA 200.8	374793		
70326093006	D226 HEALTH CENTER CLINIC 2	EPA 200.8	374793		
70326093007	D227 HEALTH CENTER CLINIC 2	EPA 200.8	374793		
70326093008	D242 KITCHEN	EPA 200.8	374793		
70326093009	D242 F KITCHEN	EPA 200.8	374793		


REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.

WO#: 70326093



70326093



Pace
 Pace Analytical Long Island NY
 575 Broad Hollow Rd, Melville, NY 11747

Company Name: NYS School for the Blind
Street Address: 2A Richmond Avenue
 Batavia, NY 14020

Contact/Report To: Don Virgilio
Phone #: 585-343-5384
E-Mail: donald.virgilio@nysed.gov
Cc E-Mail:

Invoice To: Marie Cone
Invoice E-Mail: marie.cone@nysed.gov
Purchase Order # (if applicable):

Quote #:

County / State origin of sample(s): New York

Regulatory Program (DW, RCRA, etc.) as applicable: Reportable: Yes No

Time Zone Collected: AK PT MT CT ET

Data Deliverables:

Level II Level III Level IV

EQUIS

Other

Matrix Codes (insert in Matrix box below): Drinking Water (DW), Ground Water (GW), Waste Water (WW), Product (P), Soil/Solid (SS), Oil (OL), Wipe (WP), Tissue (TS), Bioassay (B), Vapor (V), Surface Water (SW), Sediment (SED), Sludge (SL), Caulk (CK), Leachate (LL), Blossid (BS), Other (OT)

Customer Project #: LEAD TESTING

Site Collection Info/Facility ID (as applicable):

Time Zone Collected: AK PT MT CT ET

Data Deliverables:

Level II Level III Level IV

EQUIS

Other

Matrix Codes (insert in Matrix box below): Drinking Water (DW), Ground Water (GW), Waste Water (WW), Product (P), Soil/Solid (SS), Oil (OL), Wipe (WP), Tissue (TS), Bioassay (B), Vapor (V), Surface Water (SW), Sediment (SED), Sludge (SL), Caulk (CK), Leachate (LL), Blossid (BS), Other (OT)

Customer Sample ID	Matrix *	Composite Start		# Cont.	Res. Chlorine Results	Units
		Date	Time			
25	DW	11-16-24	6:11am			
29	DW	11-16-24	6:14am			
33	DW	11-16-24	6:17am			
54	DW	11-29-24	6:15am			
80	DW	11-29-24	6:24am			
61	DW	11-29-24	6:20am			
98	DW	11-29-24	6:22am			
99	DW	11-29-24	6:26am			
99	DW	11-29-24	6:27am			

Additional Instructions from Pace*:

Collected By: Grant Bunch
(Printed Name)
Signature: *Grant Bunch*

Received by/Company (Signature): *Grant Bunch*
Date/Time: 12/15/24 10:50

Received by/Company (Signature): *Marie Cone*
Date/Time: 12/16/24 4:00

Received by/Company (Signature): *Marie Cone*
Date/Time: 12/16/24 8:00

Received by/Company (Signature): *Marie Cone*
Date/Time: 12/16/24 8:00

Thermometer ID: T211
Obs. Temp (°C): 20
Corrected Factor (°C): 0
Completed Temp. (°C): 20
On Ice:

Tracking Number: 1050

Delivered by: In-Person Courier
 FedEx UPS Other

Page: 12 of 14

Customer Remarks / Special Conditions / Possible Hazards:

Lead in Drinking Water by 200.8

Preservation non-conformance identified for

Prof. Mgr: Daniel Bonitto
 AcctNum / Client ID:
 Table #:
 Profile / Template: 8697
 Preleg / Bottle Ord. ID: EZ 3-158747
 Sample Comment

*** Container Size: (1) 1L, (2) 500mL, (3) 250mL, (4) 125mL, (5) 100mL, (6) 40mL, (7) EnCore, (8) TerraCore, (9) 90mL, (10) Other

*** Preservative Types: (1) None, (2) HNO3, (3) H2SO4, (4) HCl, (5) NaOH, (6) Zn Acetate, (7) NaHSO4, (8) Sod. Thiosulfate, (9) Ascorbic Acid, (10) MeOH, (11) Other

Specify Container Size **

Identify Container Preservative Type ***

Analysis Requested

80017

Profile #:

Client: **BLIND**

COC Page

Work ID: **Lead Testing 1/16**

of

Add SCLOGFD to first sample for field charge

Use Point Number Spreadsheet

Multiday Project

COC Line Item	Matrix	Container	Coefficient	CG9U	VG9C	VG9H	VG9S	VG9T	DG9Y	DG9P	DG9A	DG6T	DG9S	AG4U	AG3U	AG2U	AG1U	AG1H	AG1A	AG5U	AG4A	CG1U	WG90	WG40	BP4U	BP3U	BP2U	BP1U	BP3S	BP2S	BP4N	BP3N	BP2N	BP3C	BP3T	BP3S	BP3R	BP1Z	BP1N	BP1B	SP5T	WG2U	WG3U	WG4U	WG5U	WG6U	WG7U	ZPLC	GN	WP	TEDL	BG1H	BG1N	IOC	SOC
---------------	--------	-----------	-------------	------	------	------	------	------	------	------	------	------	------	------	------	------	------	------	------	------	------	------	------	------	------	------	------	------	------	------	------	------	------	------	------	------	------	------	------	------	------	------	------	------	------	------	------	------	----	----	------	------	------	-----	-----

Container Coeffs

	Glass	Plastic
VG9U	40mL unpres clear vial	125mL unpreserved plastic
VG9C	40mL Ascorbic-HCl clear vial	250mL unpreserved plastic
VG9H	40mL HCl clear vial	500mL unpreserved plastic
VG9S	40mL Sulfuric clear vial	1 liter unpreserved plastic
VG9T	40mL Na Thiosulfate vial	1 liter unpreserved plastic
DG8Y	40mL Citrate-Na Thiosulfate	250mL HNO3 plastic
DG8P	40mL amber vial - TSP	250mL HNO3 plastic
DG9A	Ascorbic/Maleic Acid 40mL	500mL HNO3 plastic
DG6T	Na Thio 60mL Vial	250mL H2SO4 plastic
DG9S	Ammonium Cl/CuSO4 40mL	500mL H2SO4 plastic
CG1U	1L Unpres Jar (Con Eco)	NaOH 250mL bottle
WG90	8oz clear soil jar	250mL Trizma
AG4U	1L Ammonium Chloride	250mL Ammonium Acetate
AG5U	100mL unpres Amber Glass	250mL NH4SO4-NH4OH
AG4A	Ammonium Cl 120mL bottle	1L NaOH, Zn Acetate
BP1N	1L HNO3 plastic	BP1N 1L HNO3 plastic
BP1B	Na Thiosulfate Amber Bottle	BP1B Na Thiosulfate Amber Bottle

	Misc
SP5T	120mL Collform Na Thio
R	Terracore Kit
WG2U	12oz Unpreserved Jar
WG3U	4oz Unpreserved Jar
WG4U	8oz Unpreserved Jar
WG5U	16oz Unpreserved Jar
ZPLC	Ziplock Bag
TEDL	Tedlar Bag
BG1H	1L HCL Clear Glass
GN	General
WP	Wipe
LLHG	Low Level Hg Bottles
BG1N	1L HNO3 Clear Glass

	IOC
BP1U	1L unpreserved plastic
BP3N*	250mL HNO3 plastic
BP3C	250mL Sodium Hydroxide
AG3U	500mL unpres amber glass
BP3U	250mL unpreserved plastic

* Can also be a BP4N

	SOC
VG9T	40mL Na Thio amber vial
DG9A	40mL Ascorbic acid/maleic acid vials
DG9Y	Citrate/Na Thiosulfate 40mL
DG6T	Na Thiosulfate 60mL vial
DG6M	MonoChloric/Na Thio 60mL
AG3U	1250mL unpres amber glass
AG3T	1Na Thiosulfate 250mL bottle
BP1B	1Na Thiosulfate Amber bottle
AG1T	1Na Thiosulfate 1L Amber
AG1A	1525.3 Chemical Blend

	Matrix
WT	Water
SL	Solid
NAL	Non-aqueous Liquid
OL	OIL
WP	Wipe
DW	Drinking Water

Sender initials

Additional Comments

WO# : 70326093
PM: DHB Due Date: 12/13/24
CLIENT: BLIND

WO#: 70326093

Client Name: BLIND

Project #

PM: DHB

Due Date: 12/13/24

Courier: Fed Ex UPS USPS Client Commercial Parcel Other

CLIENT: BLIND

Tracking #:

Custody Seal on Cooler/Box Present: Yes No Seals intact: Yes No Temperature Blank Present: Yes No
Packing Material: Bubble Wrap Bubble Bags Ziploc None Other Type of Ice: Wet Blue None

Thermometer Used: THZU Correction Factor: 0 Samples on ice, cooling process has begun
Cooler Temperature ("C): 20 Cooler Temperature Corrected ("C): 2.0 Date/Time 5035A kits placed in freezer _____

Temp should be above freezing to 6.0°C

USDA Regulated Soil N/A, water sample

Did samples originate in a quarantine zone within the United States: AL, AR, CA, FL, GA, ID, LA, MS, NC, NM, NY, OK, OR, SC, TN, TX, or VA (check map)? Yes No

Did samples originate from a foreign source including Hawaii and Puerto Rico)? Yes No

If Yes to either question, fill out a Regulated Soil Checklist (ENV-FRM-MELV-0076) and include with SCUR/COC paperwork.

Date and Initials of person examining contents: 12/9/24 CJ

	COMMENTS:
Chain of Custody Present: <input checked="" type="checkbox"/> Yes <input type="checkbox"/> No	1.
Chain of Custody Filled Out: <input checked="" type="checkbox"/> Yes <input type="checkbox"/> No	2.
Chain of Custody Relinquished: <input checked="" type="checkbox"/> Yes <input type="checkbox"/> No	3.
Sampler Name & Signature on COC: <input checked="" type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A	4.
Samples Arrived within Hold Time: <input checked="" type="checkbox"/> Yes <input type="checkbox"/> No	5.
Short Hold Time Analysis (<72hr): <input checked="" type="checkbox"/> Yes <input type="checkbox"/> No	6.
Rush Turn Around Time Requested: <input checked="" type="checkbox"/> Yes <input type="checkbox"/> No	7.
Sufficient Volume: (Triple volume provided for MS/MSD) <input checked="" type="checkbox"/> Yes <input type="checkbox"/> No	8.
Correct Containers Used: <input checked="" type="checkbox"/> Yes <input type="checkbox"/> No	9.
-Pace Containers Used: <input checked="" type="checkbox"/> Yes <input type="checkbox"/> No	
Containers Intact: <input checked="" type="checkbox"/> Yes <input type="checkbox"/> No	10.
Filtered volume received for Dissolved tests: <input type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A	11. Note: if sediment is visible in the dissolved container.
Sample Labels match COC: <input checked="" type="checkbox"/> Yes <input type="checkbox"/> No	12.
-Includes date/time/ID/Analysis Matrix: <input checked="" type="checkbox"/> SL <input type="checkbox"/> WT <input type="checkbox"/> OIL <input type="checkbox"/> OTHER	

Date and Initials of person checking preservation: 12/9/24 CJ

All containers needing preservation have been: <input checked="" type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A	13. <input type="checkbox"/> HNO ₃ <input type="checkbox"/> H ₂ SO ₄ <input type="checkbox"/> NaOH <input type="checkbox"/> HCl
pH paper Lot # <u>205324</u>	Sample #
All containers needing preservation are found to be in compliance with method recommendation? (HNO ₃ , H ₂ SO ₄ , HCl, NaOH > 9 Sulfide, <input checked="" type="checkbox"/> Yes <input type="checkbox"/> No <input type="checkbox"/> N/A)	
NAOH > 12 Cyanide	
Exceptions: VOA, Coliform, TOC/DOC, Oil and Grease, DRO/8015 (water).	Initial when completed: Lot # of added preservative: Date/Time preservative added:
Per Method, VOA pH is checked after analysis	
Samples checked for dechlorination: <input type="checkbox"/> Yes <input type="checkbox"/> No <input checked="" type="checkbox"/> N/A	14. Positive for Res. Chlorine? Y N
KI starch test strips Lot #	
Residual chlorine strips Lot #	15. Positive for Sulfide? Y N
SM 4500 CN samples checked for sul: <input type="checkbox"/> Yes <input type="checkbox"/> No <input checked="" type="checkbox"/> N/A	
Lead Acetate Strips Lot #	
Headspace in ALK Bottle (>6mm): <input type="checkbox"/> Yes <input type="checkbox"/> No <input checked="" type="checkbox"/> N/A	16.
Headspace in VOA Vials (>6mm): <input type="checkbox"/> Yes <input type="checkbox"/> No <input checked="" type="checkbox"/> N/A	17.
Trip Blank Present: <input type="checkbox"/> Yes <input type="checkbox"/> No <input checked="" type="checkbox"/> N/A	
Trip Blank Custody Seals Present: <input type="checkbox"/> Yes <input type="checkbox"/> No <input checked="" type="checkbox"/> N/A	

Client Notification/ Resolution:

Field Data Required? Y / N

Person Contacted: _____

Date/Time: _____

Comments/ Resolution: _____

* PM (Project Manager) review (which includes the SCUR) is documented electronically in LIMS.